



Certificate No. : L2479-141110

財團法人全國認證基金會  
Taiwan Accreditation Foundation

## Certificate of Accreditation

This is to certify that

**Fanforce Corp.**

**Fans Laboratory**

No.63-8, Yongcheng Rd., Taiping Dist., Taichung City, Taiwan (R.O.C.)

**is accredited in respect of laboratory**

**Accreditation Criteria** : ISO/IEC 17025: 2005  
**Accreditation Number** : 2479  
**Originally Accredited** : December 02, 2011  
**Effective Period** : December 02, 2014 to December 01, 2017  
**Accredited Scope** : Testing Field, see described in the Appendix

Jay-San Chen  
President, Taiwan Accreditation Foundation  
Date : November 10, 2014



Certificate No. : L2479-141110

財團法人全國認證基金會  
Taiwan Accreditation Foundation

Accreditation Number : 2479

Laboratory Head : CHANG, Shih-Ta

▀ 16.13 Machine and Equipment

Fans and Blowers

M996 Fan Performance Test

ANSI/AMCA 210 Fig. 12, Fig. 15

ISO 5801 figure 40(e), figure 41

Air Flow Rate: 0 m<sup>3</sup>/min to 620 m<sup>3</sup>/min

Static Pressure: 0 Pa to 29400 Pa (0 mmAq to 3000 mmAq)

Approval Signatory: FU, Wen-Chang; CHANG, Shih-Ta; CHEN, Jhih-Ping

(Null below)